



620: UPHOLSTERED BLOOD-DRAW / PHLEBOTOMY CHAIR WITH PADDED ARMS - ASSEMBLY AND OPERATION INSTRUCTIONS

SAVE THESE INSTRUCTIONS FOR FUTURE USE

- ⚠ **WARNING: Important! Read and understand these instructions before assembling or using the Blood-Draw / Phlebotomy Chair. If you do not understand any part of these warnings, cautions or instructions, contact a healthcare professional for direction in the use of this product. If the Blood-Draw / Phlebotomy Chair is not properly assembled and used, personal injury and damage to the Blood-Draw / Phlebotomy Chair could result.**
- ⚠ **WARNING: If components are damaged or missing, contact your dealer immediately. DO NOT use substitute parts. Use only Intensa replacement parts. Non-Intensa replacement parts could cause personal injury and damage to the Blood-Draw / Phlebotomy Chair.**
- ⚠ **WARNING: This product can expose you to chemicals including Di(2-ethylhexyl)phthalate (DEHP) which is known to the State of California to cause cancer and birth defects or other reproductive harm. For more information go to www.p65warnings.ca.gov/furniture.**
- ⚠ **WARNING: 620 Chair maximum weight capacity is 800 lb (363 kg), EVENLY DISTRIBUTED.**
- ⚠ **WARNING: Intensa, Inc. assumes no responsibility for any damage or injury caused by improper installation or use of this product.**

INTENDED USE

The Blood-Draw / Phlebotomy Chair is intended as a medical-room chair where a patient can sit comfortably to allow medical personnel to draw blood.



620 Blood-Draw / Phlebotomy Chair

UNPACKING

Check for any obvious damage to the carton or contents (after unpacking). If damage is evident, immediately notify the carrier and Intensa.

Open the box and carefully remove the assembled frame and other components. Inside the box are:

- 1) One frame with seat and arm receivers attached and arms installed
- 2) One back
- 3) Four washers and four #10 x 2" Phillips flat head wood screws (to attach back to frame)
- 4) One Wireform Wallsaver, if ordered

CHAIR ASSEMBLY

Tools required: *Phillips screwdriver*

Attaching Back to Frame



1. Place the flat side of the back in position against the rail at rear of frame with curve upward, as shown at above left, centering the back evenly against the chair frame. Use provided wood screws to secure with back frame.
2. Secure the back to the frame, as shown above. Use a Phillips screwdriver and the provided four screws and washers, two screws and washers per side as shown, until tight.

Arm Height and Position Adjustment

1. Pull the plunger-knob, located on front of arm receiver tube, out to release the arm.
2. Slide the arm to the desired height and position.
3. Release the plunger-knob to lock the arm at desired height and position.

⚠ WARNING: Ensure arms are locked securely in place before proceeding.

Wireform Wallsaver Installation, if Ordered



Wallsaver End & Chair Leg Slot



Wallsaver Shown From Back After Installation

1. Position the Wireform Wallsaver with bent sections facing down and tab ends facing inward as shown at above left.
2. Gently squeeze both Wallsaver tab ends inward toward the center of the Wallsaver and hold the tabs parallel to the slots located at the bottom of the chair's back legs.
3. Slide the tabs into the base holes and let the Wallsaver gently expand outward to lock in place.

Wireform Wallsaver Removal

To remove the Wallsaver, squeeze the ends toward the center of the Wallsaver until the end tabs slide out of the frame base holes.

OPERATION

- ⚠ **WARNING: Ensure the Blood-Draw / Phlebotomy Chair is securely assembled as previously described before each use. Monitor patient to ensure Blood-Draw / Phlebotomy Chair does not adversely affect patient in any way. If so, discontinue use immediately.**

CARE AND MAINTENANCE

Cleaning and Disinfection

- ⚠ **WARNING: Check the Blood-Draw / Phlebotomy Chair before each use for signs of damage. If damage is found, DO NOT use the Blood-Draw / Phlebotomy Chair. Contact Intensa immediately.**

To clean Blood-Draw / Phlebotomy Chair, hand wash gently with warm water and a mild soap or detergent, rinse with clear water, and air dry.

To disinfect Blood-Draw / Phlebotomy Chair, use a diluted solution of 10% (or less) household bleach containing 5.25% sodium hypochlorite, rinse with clear water, and air dry.

- ▲ **NOTICE: DO NOT use abrasive cleaners or scouring pads on any part of the Blood-Draw / Phlebotomy Chair.**

- ⚠ **WARNING: Inspect all components periodically and replace as needed.**

Every three months

- Check back to ensure screws are tight; if not, tighten.

Upholstery maintenance and stain removal

- Ordinary dirt: wash with warm water and a mild soap or detergent.
- Stubborn dirt: using a soft bristle brush, wash with warm water and a mild soap or detergent.
- Ground-in dirt: scrub with powdered cleanser or similar detergent.
- Chewing gum: scrape carefully; remove with kerosene or naphtha.
- Nail polish and remover: blot immediately to prevent damage.
- Ball point pen: blot immediately with a white cloth dampened in water or mineral oil.
- Paint, shoe polish, heel marks: remove immediately! Use a white cloth dampened in kerosene, naphtha or turpentine. Do not use paint remover or liquid brush cleaners.
- Tars, asphalt, creosote: remove to prevent staining. Clean area with kerosene or naphtha.

- ⚠ **WARNING: When using kerosene or naphtha, do so in a well ventilated area and keep it away from fire, flame, and chair occupant.**

- Waxing and refinishing: improves wearability. Use any quality paste wax.

LIMITED WARRANTY

SCOPE OF WARRANTY

Intensa, Inc. ("Intensa") warrants to the original purchaser only that it will replace or repair components, at Intensa's sole discretion, that are defective in material or workmanship under normal use and service. All warranties are conditioned upon the proper use of the products strictly in accordance with good commercial practice and applicable Intensa instructions and manuals, including proper use and maintenance. To the extent that a component is warranted by a third party, Intensa conveys all of its rights under that warranty to the original purchaser, to the extent permitted. This limited warranty shall only apply to defects that are reported to Intensa's customer service team within the applicable warranty period and which, upon examination by Intensa or its authorized representative, prove to be a warranty item. This limited warranty is not transferable. Within the guidelines set forth in this document, the following components are warranted for the time period set forth below:

Components *

Frame:.....	5 years
Mechanical Components:	3 years
Original and Replacement Upholstered Tops †:	1 year
Replacement Parts ‡:.....	90 days

* Labor is not included in the warranty.

† Upholstery is only warranted on material supplied by Intensa.

‡ The warranty period is as designated above. If a part is replaced under warranty, the original warranty period will not be affected. All other replacement parts will be subject to the warranty period specified.

The applicable warranty period shall commence from date of shipment to the original customer, unless there is an expiration date on the component in which case the warranty shall expire on the earlier of warranty period or the expiration date.

OBTAINING WARRANTY SERVICE

This limited warranty shall only apply to defects that are reported to the Distributor from whom the Customer purchased the product within the applicable warranty period. If there is not a Distributor, you must contact Intensa directly by calling 1.336.884.4096, or by e-mailing a request to intensa@grahamfield.com. Specific directions will be provided by the Customer Service Representative. Failure to abide by the specific directions will result in denial of the warranty claim.

EXCLUSIONS

The warranty does not cover and Intensa shall not be liable for the following:

- 1) Defects, damage, or other conditions caused, in whole or in part, by misuse, abuse, negligence, alteration, accident, freight damage, tampering or failure to seek and obtain repair or replacement in a timely manner;
- 2) Products which are not installed, used, or properly cleaned and maintained as required in the official manual for the applicable product;
- 3) Products considered to be of a non-durable nature including, but not limited to: tips, casters, filters, fuses, gaskets, lubricants, and charts;
- 4) Accessories or parts not provided by Intensa;
- 5) Matching of color, grain or texture except to commercially acceptable standards;
- 6) Changes in color caused by natural or artificial light;
- 7) Charges by anyone for adjustments, repairs, replacement parts, installation or other work performed upon or in connection with such products which are not expressly authorized in writing, in advance, by Intensa;
- 8) Any labor or shipping charges incurred in the replacement part installation or repair;
- 9) Costs and expenses of regular maintenance and cleaning; and
- 10) Representations and warranties made by any person or entity other than Intensa.

ENTIRE WARRANTY, EXCLUSIVE REMEDY AND CONSEQUENTIAL DAMAGES DISCLAIMER

THIS WARRANTY IS INTENSA'S ONLY WARRANTY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED. INTENSA MAKES NO IMPLIED WARRANTIES OF ANY KIND INCLUDING ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. IF ANY MODEL OR SAMPLE WAS SHOWN TO THE CUSTOMER, SUCH MODEL OR SAMPLE WAS USED MERELY TO ILLUSTRATE THE GENERAL TYPE AND QUALITY OF THE PRODUCT AND NOT TO REPRESENT THAT THE PRODUCT WOULD NECESSARILY CONFORM TO THE MODEL OR SAMPLE IN ALL RESPECTS. THIS WARRANTY IS LIMITED TO THE REPAIR OR REPLACEMENT OF THE DEFECTIVE PARTS. INTENSA SHALL NOT BE LIABLE FOR AND HEREBY DISCLAIMS ANY DIRECT, SPECIAL, INDIRECT, INCIDENTAL, EXEMPLARY OR CONSEQUENTIAL DAMAGES, INCLUDING, BUT NOT LIMITED TO: DAMAGES FOR LOSS OF PROFITS OR INCOME, LOSS OF USE, DOWNTIME, COVER, OR EMPLOYEE OR INDEPENDENT CONTRACTOR WAGES, PAYMENTS AND BENEFITS. CERTAIN STATES MAY CONFER ADDITIONAL RIGHTS REGARDING WARRANTIES AND IN THOSE STATES INTENSA'S LIABILITY AND THE LIABILITY OF INTENSA'S SUPPLIERS, SHALL BE LIMITED TO THE FULLEST EXTENT PERMITTED BY LAW.

The warranties contained herein, together with Intensa's current Terms and Conditions, contain all the representations and warranties with respect to the subject matter of this document, and supersede all prior negotiations, agreements and understandings with respect thereto. The recipient of this document hereby acknowledges and represents that it has not relied on any representation, assertion, guarantee, warranty, collateral contract or other assurance, except those set out in this document.

For additional information on these products or this warranty, please contact an Intensa Customer Service Representative.

NOTES:

- 1) Additional terms and conditions may apply. See Intensa's General Terms and Conditions on its website and the specific warranties, which may accompany the specific product.
- 2) Freight claims must be notated on the appropriate shipping documents and must be made with immediacy. International, federal and state regulations govern specific requirements for freight claims. Failure to abide by those regulations may result in a denial of the freight claim. Intensa will assist you in filing the freight claim.
- 3) Claims for any short shipment must be made within three (3) days of the invoice date.



© 2022, Intensa, Inc. All Rights Reserved.
Intensa is a Wholly Owned Subsidiary of GF Health Products, Inc.
GF Health Products, Inc. is an ISO 13485:2016 / MDSAP Certified Company
Manufactured by: Intensa, Inc., 1810 S. Elm St., High Point, NC 27260, 1.336.884.4096
intensa@grahamfield.com
intensa.net
Made in U.S.A.